

Specifications

Name: Chlorobenzene
Catalog No: CS1311-002
Grade: HPLC/Spectro

Test Name

Description

Clear colorless liquid free of suspended matter

Assay (by GC-FID)

(Corrected for Water)

Color

Residue after Evaporation

Titration Acid

Water

(by KF coulometry)

Optical Absorbance at

288 nm

300 nm

325 nm

350 nm

400 nm

Limits

Passes Test

Min 99.9 %

Max 30 APHA

Max 3 ppm

Max 0.004 meq/g

Max 0.03 %

Max 1.00 AU

Max 0.06 AU

Max 0.04 AU

Max 0.02 AU

Max 0.01 AU

4/16/2024

Form Number : FN-012

Revision Date: 06-01-2016

Page 1 of 1



Clear focus. Consistent results. Complete confidence.