

Specifications

Name: Methyl Alcohol
Catalog No: MP1924-002
Grade: Purge & Trap

<u>Test Name</u>	<u>Limits</u>
Water (by KF coulometry)	Max 0.200 %
Refractive Index at 20°C	1.3280-1.3286
Residue after Evaporation	Max 1 mg/L
Assay (by GC-FID)	Min 99.9 %
All Aromatics	Passes Test
Ketones	
(2-Butanone)	Max 100 ng/ml
(Acetone)	Max 50 ng/ml
(Methyl Isoamyl Ketone)	Max 50 ng/ml
(All other detectable ketones)	Max 50 ng/ml
Chlorohydrocarbons (non-aromatic)	
(Dichloromethane)	Max 50 ng/ml
(Chloroform)	Max 50 ng/ml
(Carbon Tetrachloride)	Max 50 ng/ml
(All other detectable chlorohydrocarbons)	Max 50 ng/ml
Alkanes	
(n-Pentane)	Max 50 ng/ml
(n-Hexane)	Max 50 ng/ml
(n-Heptane)	Max 50 ng/ml
(2,2,4-Trimethylpentane)	Max 50 ng/ml
(All other detectable alkanes)	Max 50 ng/ml
All other detectable peaks (as n-Hexane)	Max 50 ng/ml