

Specifications

Name: Water
Catalog No: WS2211-048
Grade: HPLC

Test Name

Description

Clear colorless liquid free of suspended matter

Color

Fluorescence at emission maximum for Impurities at

450 nm

LC gradient elution suitability at

254 nm

Residue after Evaporation

Specific Conductance at 25°C

(at time of packaging)

Substances Reducing Permanganate

Optical Absorbance at

190 nm

200 nm

210 nm

250 nm

400 nm

Limits

Passes Test

Max 5 APHA

Max 0.1 ppb

Passes Test

Max 1 ppm

Max 1.5 μ S/cm

Passes Test

Max 0.01 AU

Max 0.01 AU

Max 0.01 AU

Max 0.005 AU

Max 0.005 AU

10/5/2024

Form Number : FN-012

Revision Date: 06-01-2016

Page 1 of 1



Clear focus. Consistent results. Complete confidence.