

## Certificate Of Analysis

January 18, 2010

Name: Tetrahydrofuran (Inhibitor Free)  
Catalog No: TS2121  
Grade: HPLC/Spectro  
Tedia Lot No. 1001025  
Expiry Date: January 13, 2011

This is to certify that this lot was tested and found to comply with the specifications for the product. The following are the actual analytical results obtained.

<u>Test Name</u>	<u>Limits</u>	<u>Actual</u>
Assay (by GC-FID)	Min 99.8 %	
Color	Max 10 APHA	<
Peroxides		
(as Hydrogen peroxide at time of packaging)	Max 0.015 %	<
Residue after evaporation	Max 5 ppm	<
Water		
(by KF coulometry)	Max 0.03 %	
Optical absorbance at		
212 nm	Max 1.00	
225 nm	Max 0.50	
250 nm	Max 0.17	
350 nm	Max 0.01	<
Titrate acid	Max 0.0008 meq/g	



Sherwin M. Kane  
Chemist

Tedia Company, Inc. certifies that this product was derived from synthetic raw materials and the manufacturing process excluded contamination with any animal products.